



EC-Certificate

(full quality assurance system) according to annex II (excluding section 4) of Medical Devices Directive 93/42/EEC

It is herewith confirmed by

BSI Group Deutschland GmbH

Eastgate, Hanauer Landstrasse 115 60314 Frankfurt am Main Germany

in its function as Notified Body (0535), that the manufacturer:



NSK WT GmbH Portitzer Straße 69d 04425 Taucha Germany

concerning the medical devices

Disinfector

UMDNS: 11-278

(products/variants specified in appendix)

fulfils the requirements according to Annex II (excluding section 4) Medical Devices Directive 93/42/EEC. The manufacturer has established a quality assurance system for the design, production and final inspection of the specified devices.

For the placing on the market of class III products an additional Annex II section 4 certificate is required.

The appendix is part of this certificate and contains 1 page.

Report No.: SMO7957250/SMO7957255

Certificate No.: CE 595189

Current Issue Date: October 21, 2013

Certification Body



Zentralstelle der Länder für Sicherheitstechnik ZLS-NB-67/12

First Issue Date: October 21, 2013.

Based on periodical surveillance this certificate is valid until July 16, 2018.





Appendix of EC-Certificate

(full quality assurance system)

according to annex II (excluding section 4) of Medical Devices Directive 93/42/EEC

Certificate No.: CE 595189

Medical devices of the manufacturer:



NSK WT GmbH Portitzer Straße 69d 04425 Taucha Germany

Name of product	Variant	Item	UMDNS/ GMDN	Class
iCare+	C3	S103001	11-278	IIb

ZLS Notified by
Zentralstelle der Länder
für Sicherheitstechnik
ZLS-NB-67/12

Certification Body

Frankfurt am Main, October 21, 2013